



Schematic arrangement of BSK12 and location in purpose designed trough with conical profiled lid. Approximate dimensions are in mm.



**Scientific
Systems
Design Inc**

1365 #14 Mid-Way Blvd
Mississauga
Ontario L5T 2J5
CANADA

Phone: 1 905 608 9307
ssd@scisys.info
www.scisys.info

Innovative Engineering for Science

BRAIN SLICE KEEPER

BSK12

BRAIN SLICE KEEPER

CAUTION !

YOUR BRAIN SLICE KEEPER IS A PRECISION ENGINEERED TOOL FOR SCIENTIFIC RESEARCH. PLEASE TAKE A FEW MINUTES TO FAMILIARISE YOURSELF WITH THE KEEPER AND READ THROUGH THIS SHORT MANUAL BEFORE ATTEMPTING TO USE IT.

DO NOT USE ALCOHOL OR SIMILAR SOLVENTS IN ANY CONCENTRATION ON ANY PART OF THE KEEPER SINCE AS WITH MOST ACRYLICS, IT MAY FRAGMENT OR DEVELOP HAIR-LINE CRACKS. DO NOT AUTOCLAVE AS HEAT CYCLING CAN CAUSE STRESS CRACKS IN ACRYLIC.

OPERATION

The BSK12 Brain Slice Keeper has been designed to pre-incubate a large number of brain slices prior to transfer into recording chambers or for the incubation of preparations in experimental test media. This keeper consists of twelve closely fitting acrylic 'C' rings that fit into their own through-hole, located in a circular manifold. Removable netting is wedged between the 'C' ring and it's hole.

FEATURES

- * Slices maintained for many hours in incubating media
- * Twelve separate rings allow separation of different types of preparations
- * Slices supported on a quick change nylon net
- * Modular design simple to set up and maintain

In operation, the twelve rings located on a manifold with netting are totally immersed in an ACSF-filled purpose designed holding vessel into which a 95% O₂, 5% CO₂ gas mixture is supplied through an airstone bubbler. The bubbler is located in a central tube chamber connected to the upper manifold. The bubbles rising from the base saturate the ACSF and provide constant circulation of medium to the slices which rest on the nylon nets. Bubbles are restricted to rising in the central chamber and are prevented from being trapped under the slices, thereby ensuring a continuous circulation of medium. Slices remain viable for many hours in these conditions. The BSK12 together with the holding vessel can be easily placed into a water bath for regulating the incubating temperature as desired. A lid with a conical profile ensures that condensation drops do not fall directly above the slices. When in place, the fluid level is maintained at least 3mm above the 'C' rings. This ensures circulation of fluid over the top and downwards towards the slices. The typical fluid volume with BSK12 is 500ml but can be reduced by adding large glass spheres into the base.

TEMPERATURE CONTROL

Temperature can be maintained by submerging the trough into a water bath set to the desired temperature. The heated water bath fill level should be about 15mm below the fill level of the BSK12 to prevent floating and instability. Bubbling within the BSK12 ensures circulation and uniform temperature of the incubating media around the slices.

MAINTENANCE

Alcohol should never be used on the acrylic manifold and rings of the slice keeper for cleaning purposes even at low concentrations because it dehydrates and produces hair-line cracks in acrylic plastics. Only the plastic trough can be cleaned with alcohols. The acrylic component can be cleaned with special laboratory detergents such as *Micro-90™ which completely rinses out. The air stone must be removed when using such detergents. Heavy deposits of salts should be washed out with distilled water



BSK12 Brain slice keeper with rings and net in place and with lid in place (inset). A needle valve is used to control the flow of oxygen mixture to the ceramic bubbler.

overnight and carbonate salts treated with mild acids such as citric acid. The most common contaminant is fungal growth, this can be avoided by washing out with distilled water and drying out completely at the end of each day. Hydrogen peroxide solution 30 Vols, 1/10 dilution is also an effective cleaning agent followed by overnight soaking in distilled water.